

## CERTIFICATE OF ANALYSIS

<b>Batch No:</b>	<b>22-05306</b>	<i>Page</i>	<b>Page 1 of 2</b>
<i>Final Report</i>	<b>939446</b>	<i>Laboratory</i>	Geelong Laboratory
<i>Client:</i>	<b>Chris Ryan</b> <i>For Barbara Hammond</i>	<i>Address</i>	16 Crown Street, South Geelong, VIC 3220
<i>Contact:</i>	<b>Chris Ryan</b>	<i>Phone</i>	03 5246 9400
<i>Address:</i>	85 Bimbadeen Drive Fairhaven Vic 3231	<i>Fax</i>	
		<i>Contact:</i>	Adele Vance Geelong Laboratory Manager Adele.Vance@alsglobal.com
<i>PO No:</i>	<b>Not Available</b>	<i>Date Sampled:</i>	<b>20-Jan-2022 - 21-Jan-2022</b>
<i>Sampler Name:</i>	<b>Client</b>	<i>Date Samples Received:</i>	<b>21-Jan-2022</b>
<i>ALS Program Ref:</i>	<b>SUNDRY_GEE</b>	<i>Date Issued:</i>	<b>28-Jan-2022</b>
<i>Program Description:</i>	<b>Sundry Customer Program for Geelong</b>		
<i>Client Ref:</i>	<b>Water Sampling</b>		

The hash (#) below indicates methods not covered by NATA accreditation in the performance of this service.

Analysis	Method	Laboratory	Analysis	Method	Laboratory
Colilert (200)	MM514	Geelong	EC	WA010	Geelong
MS Total Metals	WG020A	Scoresby	pH	WA005	Geelong
Turbidity	WA045	Geelong			



*Result for pH in water tested in the laboratory may be indicative only as holding time is generally not achievable.*

**Analysis Notes :**

**According to Australian Drinking Water Guidelines, Coliforms and Escherichia coli should not be detected in drinking water.** If detected, the water must be treated. For further information, please visit the Department of Health web site by clicking on the link below;  
<https://www2.health.vic.gov.au/Api/downloadmedia/%7B1C96DDEA-8B63-4020-A88E-6E106FE2DBC7%7D>

**Signatories**

Name	Title	Name	Title
<b>Adele Vance</b>	<b>Geelong Laboratory Manager</b>	<b>Adele Vance</b>	<b>Geelong Laboratory Manager</b>
<b>Mark Sheedy</b>	<b>Analyst</b>	<b>Ricky Singh</b>	<b>Analyst</b>

 <small>Accreditation No. 992 Accredited for compliance with ISO/IEC 17025 - Testing</small>	 <small>Accreditation No. 992 Accredited for compliance with ISO/IEC 17025 - Testing</small>	<p>Samples not collected by ALS and are tested as received.</p> <p>Soil microbiological testing was commenced within 4 days from the day collected unless otherwise stated.</p> <p>Water microbiological testing was commenced on the day received and within 24 hours of sampling unless otherwise stated.</p> <p>MM524: Plate count results &lt;10 per mL and &gt;300 per mL are deemed as approximate.</p> <p>MM526: Plate count results &lt;2,500 per mL and &gt;250,000 per mL are deemed as approximate.</p> <p>Calculated results are based on raw data.</p> <p>Legionella species refers to Legionella species other than Legionella pneumophila</p> <p>Measurement Uncertainties values for your compliance results are available at this link</p> <p>Samples are tested within holding time unless otherwise stated.</p>
---	--	--

The report shall not be reproduced,  
except in full and results relate only to  
the items tested.



Sample No	Site Code	Site Description	Sample Type	Sampled Date/Time
7318659		Inlet Mouth Aireys Inlet	WATER	20/01/22 15:30
7318660		Coastal Court Aireys Inlet	WATER	20/01/22 16:30
<del>7318661</del>		<del>Tank Water Bimbadeeen Dr</del>	<del>WATER</del>	<del>21/01/22 09:30</del>

Analysis - Analyte	Sample No.	7318659	7318660	<del>7318661</del>
	Site Code Units			
pH - pH, units	Units			8.3
EC - Electrical Conductivity @ 25C	uS/cm			100
Turbidity - Turbidity, NTU	NTU			0.1
MS Total Metals - Copper	mg/L			0.09
MS Total Metals - Iron	mg/L			0.02
MS Total Metals - Lead	mg/L			0.01
MS Total Metals - Zinc	mg/L			0.12
Colilert (200) - Coliforms MPN Colilert	orgs/100mL	>2400	>2400	8
Colilert (200) - E.coli MPN Colilert	orgs/100mL	1400	440	0

A blank space indicates no test performed.